



NOTES:
ALL WIRE SIZES NOT SPECIFIED ARE 22AWG. OR HEAVIER. IT IS RECOMMENDED THAT EACH GREEN GROUND SCREW, GREEN GROUND WIRE, AND GROUND SCREW TERMINAL ON CALL STATIONS (LABELED "G") BE CONNECTED TO EARTH GROUND, 18AWG. WIRE MINIMUM.



CORNELL Communications, Inc.
 sales@cornell.com 414-351-4660
 Web: www.cornell.com 414-351-4657(fax)

DATE: SEPTEMBER 22, 2008

TITLE: DUAL STATUS NC-102D / L-102 / E-104-1

SYSTEM: 4000 NURSE CALL

DRAWING # 90006-08

D4000W-2B